

Monaco计划系统 在乳腺癌治疗中的应用

鞍山市肿瘤医院

王璐



2018 ProKnow AAMD Breast 5 Targets

CERTIFICATE OF ACHIEVEMENT



proudly presented to

Wanglu

for participation in the
2018 Breast Plan Study

Awarded on June 4, 2018
with a
Rank of 1 out of 240
and a
Score of 150 out of 150

鞍山市肿瘤医院



计划要求

24 Key Metrics

Total Points 150






Scoring powered by PlanIQ™
Sun Nuclear Corporation

50 Gy (25 x 2 Gy) to the target volumes detailed in the plan metrics.

#	METRIC ID (28 Total Metrics)	MIN REQ	IDEAL	WEIGHT (150)
[01]	Volume (%) of the BREAST_PTV_EVAL covered by 47.5 (Gy)	90	97	20 points
[02]	Dose (Gy) covering 50 (%) of the BREAST_PTV_EVAL	54	52	3 points
[03]	Dose (Gy) covering 0.03 (cc) of the BREAST_PTV_EVAL	57	56	2.5 points
[04]	Volume (%) of the LUMPEC_PTV_EVAL covered by 47.5 (Gy)	95	99	20 points
[05]	Dose (Gy) covering 50 (%) of the LUMPEC_PTV_EVAL	54	52	3 points
[06]	Dose (Gy) covering 0.03 (cc) of the LUMPEC_PTV_EVAL	57	56	2.5 points
[07]	Structure(s) containing the global max dose point	BREAST_PTV_EVAL	LUMPEC_PTV_EVAL	3 points
[08]	Volume (%) of the SUPRACLAV_PTV covered by 47.5 (Gy)	90	97	15 points
[09]	Dose (Gy) covering 0.03 (cc) of the SUPRACLAV_PTV	55	53.5	2.5 points
[10]	Volume (%) of the AXILLARY_PTV covered by 47.5 (Gy)	90	97	15 points
[11]	Dose (Gy) covering 0.03 (cc) of the AXILLARY_PTV	55	53.5	2.5 points
[12]	Volume (%) of the IMN_PTV covered by 47.5 (Gy)	90	97	15 points
[13]	Dose (Gy) covering 0.03 (cc) of the IMN_PTV	55	53.5	2 points
[14]	Dose (Gy) covering 5 (%) of the BREAST_CONTRA	3	2	5 points
[15]	Dose (Gy) covering 0.03 (cc) of the BREAST_CONTRA	7	5	3 points
[16]	Dose (Gy) covering 5 (%) of the HEART	20	5	2 points
[17]	Dose (Gy) covering 30 (%) of the HEART	10	5	5 points
[18]	Mean dose (Gy) to the HEART	3	2	5 points
[19]	Volume (%) of the LUNG_IPSI covered by 20 (Gy)	37	20	5 points
[20]	Volume (%) of the LUNG_IPSI covered by 10 (Gy)	55	30	5 points



靶区及主要OAR的空间分布





射野角度及床角的选择

The screenshot displays a 3D visualization of a patient's head and neck region, with various anatomical structures and target volumes outlined in different colors. The interface includes a 'Structure Visibility' panel on the right, a 'Beams' panel at the bottom, and a 'Geometry' tab selected in the bottom right corner.

Structure Visibility Panel:

- Structure
- Axillary_CTV
- Axillary_PTV
- Body
- BPTV-LPTV
- Breast_Contra
- Breast_CTV
- Breast_Ipsi
- Breast_PTV
- Breast_PTV_EVAL
- Carbon Fiber
- Esophagus
- Foam Core
- Heart
- Humeral_Head_LT
- IMN_CTV
- IMN_PTV
- Lumpect_PTV_EVAL
- Lumpectomy_CTV
- Lumpectomy_GTV
- Lumpectomy_PTV
- Lung_Contra
- Lung_Ipsi
- Lung_Total
- PET Pre-Op
- PTV_ALL
- Skin
- Spinal Cord
- Supraclav_CTV
- Supraclav_PTV
- Thyroid
- Wire

Beams Panel:

4 New beam

Beam	Description	SSD (cm)	Dir	Gantry Start (deg)	Arc	Inc	Collimator (deg)	Couch (deg)	Field	Margin (cm)	Asym	Width1 (cm)	Width2 (cm)	Length1 (cm)	Length2 (cm)
4	New beam	88.40	CW	300.0	200.0	10.0	325.0	20.0	[Auto]	0.20	☒	X1 15.60	X2 14.10	Y2 13.50	Y1 13.50

<click to add a new beam>



计划参数设计

Sequencing Parameters: VMAT [X]

☒ Segment Shape Optimization

☒ Pilot Beamlets

Max Number of Arcs:

Max. # of Control Points Per Arc:

Target Dose Rate (MU/min):

Min. Segment Width (cm):

Fluence Smoothing:

☐ Constant Dose Rate

OK Cancel

Calculation Properties [X]

Grid Settings

Grid Spacing (cm):

Calculate Dose Deposition to:

Force entire volume to be treated as water: ☐

Grid Settings changes will be applied to ALL Rx IDs.

Algorithm Settings

Algorithm: Monte Carlo Photon Photon

Statistical Uncertainty (%):

☐ Per Control Point ☒ Per Calculation

OK Cancel

IMRT Prescription Parameters [X]

Minimum CT Number:
Use with Clear option.

Auto Flash Margin (cm):

Surface Margin (cm):

Beamlet Width (cm):

Target Margin:

Avoidance Margin:

Bias Contribution:

OK Cancel



一个关键词, 一个函数, 一个答疑

Fluence Smoothing

只作用于通量结果
对于通量结果的理想预期
预期越高对于二阶优化影响越大
高的预期会降低Segment数量

Margin

Beam Target Oar Surface 散射控制
Autoflash 散射补偿
Function 划分控制区域



Target EUD

它是一个线性控制函数
表现于对于靶区内冷点的罚分增益
靶区作用效果/靶区作用效果+靶区约束效果




计划函数设计

IMRT Constraints											
Pareto Constrained IMRT Parameters											
Structure	Cost Function	Enabled	Status	Manual	Weight	Reference Dose (cGy)	Multicriterial	Isoconstraint	Isoeffect	Relative Impact	
Supradlav_PTV	Target Penalty	☒	On	☒	60.00			5100.0	4904.4		
	Maximum Dose	☒	On	☐	2319.03			5248.0	5239.8	++++	
Axillary_PTV	Target Penalty	☒	On	☒	75.00			5100.0	5014.9		
	Maximum Dose	☒	On	☐	1494.05			5300.0	5304.6	++++	
	Underdose DVH	☒	On	☒	75.00	4810.0		98.00	99.65	+++	
	Target EUD	☒	On	☐	1.00			5000.0	5111.5		
	Target EUD	☒	On	☐	1.00			5000.0	5111.5		
Lung_Ipsi	Serial	☒	On	☐	0.01		☐	700.0	514.4		
	Overdose DVH	☒	On	☐	797.59	2000.0	☐	18.00	16.31	++++	
	Overdose DVH	☒	On	☐	4549.52	1000.0	☐	27.00	26.63	++++	
	Overdose DVH	☒	On	☐	445.88	500.0	☐	45.00	39.02	++++	
	Serial	☒	On	☐	0.01		☐	370.0	258.1		
	Serial	☒	On	☐	1079.54		☐	1560.0	1520.3	++++	
	Target Penalty	☒	On	☒	85.00			5100.0	4770.3		
	Maximum Dose	☒	On	☐	0.01			5505.0	5395.2		
Breast_Contra	Serial	☒	On	☐	1.47		☐	285.0	280.1	++	
	Overdose DVH	☒	On	☐	86.42	200.0	☐	4.00	3.16	+++	
	Maximum Dose	☒	On	☐	5.50			520.0	535.0	++	
	Parallel	☒	On	☐	4735.73	200.0	☐	6.50	6.07	++++	
	Maximum Dose	☒	On	☐	5.50			520.0	535.0	++	
	Parallel	☒	On	☐	4735.73	200.0	☐	6.50	6.07	++++	
Lung_Ipsi	Serial	☒	On	☐	0.01		☐	700.0	514.4		
	Overdose DVH	☒	On	☐	797.59	2000.0	☐	18.00	16.31	++++	
	Overdose DVH	☒	On	☐	4549.52	1000.0	☐	27.00	26.63	++++	
	Overdose DVH	☒	On	☐	445.88	500.0	☐	45.00	39.02	++++	
	Serial	☒	On	☐	0.01		☐	370.0	258.1		
	Serial	☒	On	☐	1079.54		☐	1560.0	1520.3	++++	
Lung_Contra	Overdose DVH	☒	On	☐	3812.03	500.0	☐	4.40	4.20	++++	
	Parallel	☒	On	☐	86.54	500.0	☐	7.50	6.51	+++	
Heart	Overdose DVH	☒	On	☐	0.01	500.0	☐	5.00	2.72		
	Serial	☒	On	☐	2295.49		☐	199.0	199.3	++++	
	Overdose DVH	☒	On	☐	0.05	200.0	☐	37.00	24.89	+	
Humeral_Head_LT	Serial	☒	On	☐	90.59		☐	1000.0	980.9	+++	
Thyroid	Serial	☒	On	☐	784.04		☐	2150.0	2098.3	++++	
	Serial	☒	On	☐	0.01		☐	3600.0	3489.1		
	Parallel	☒	On	☐	7193.01	2000.0	☐	37.50	36.45	++++	
Body	Conformality	☒	On	☐	1.60		☐	0.50	0.40	++	
	Maximum Dose	☒	On	☐	4.27			5500.0	5444.4	++	



计划剂量分布





竞赛型计划小结

- 1: 先进行摸底性条件设计，找出计划设计难点
- 2: 针对计划设计难点通过改变射野分布，治疗床角度、参数条件以及函数配置来实现目标
- 3: 最好的剂量保护是钨门遮挡或者入射规避
- 4: 纸面结果并不代表好的治疗效果



乳腺术后调强放疗

放射治疗医嘱单

放疗

姓

治疗日期:	2018 年 6 月 15 日	医师:	
计划系统:	()Pinnacle (☒)Monaco		
计划类型:	()静态调强 (☒)动态调强 ()适形		
照射部位:	左胸壁 + 左锁骨上		
FW	总剂量: (50)Gy	分割剂量: (2)Gy	次数: (25)Fx
	总剂量: ()Gy	分割剂量: ()Gy	次数: ()Fx
	总剂量: ()Gy	分割剂量: ()Gy	次数: ()Fx
结束日期:	年 月 日		

靶区	体积(%)	关系	剂量(Gy)	热点(Gy)	关系	体积(%)
FW	D95	≥	50Gy		≤	
		≥			≤	
		≥			≤	

危急器官	体积(%)	关系	剂量(Gy)	平均剂量(Gy)	最大剂量(Gy)
heart		≤		Dmean < 8Gy	
sc		≤			Dmax < 40Gy
lung	V20	≤	30%		
	V10	≤	40%		
esophagus	V35	≤	40%		
		≤			
		≤			
		≤			
		≤			
		≤			
		≤			

备注: 未见人造痰 0.3cm 鞍山市肿瘤医院




根治性乳腺癌IMRT射野分布






扫描图像决定射野布局





ISO Point 位置的选择





IMRT参数配置

IMRT Prescription Parameters

Minimum CT Number:
Use with Clear option.

-200

Auto Flash Margin (cm):

2.00

Surface Margin (cm):

0.30

Beamlet Width (cm):

0.30

Target Margin:

Narrow (3-4mm)

Avoidance Margin:

Normal (8mm)

Bias Contribution:

OK

Cancel

Sequencing Parameters: dMLC

☒ Segment Shape Optimization

☒ Pilot Beamlets

Max. # of Control Points Per Beam:

80

Target Dose Rate (MU/min):

Min. Segment Width (cm):

0.80

Fluence Smoothing:

Off

☒ Max. Sweep Efficiency

☒ Allow Move Only Segments

OK

Cancel




IMRT函数配置

IMRT Constraints										
↑ ↓ Pareto Constrained IMRT Parameters										
Structure	Cost Function	Enabled	Status	Manual	Weight	Reference Dose (cGy)	Multicriterial	Isoconstraint	Isoeffect	Relative Impact
PTV50	Quadratic Overdose	☒	On	☐	0.01	5150.0		100.0	14.5	
	Target EUD	☒	On	☒	3.00			5000.0	4887.4	
	Target Penalty	☒	On	☒	3.00			5100.0	4754.8	
lung	Parallel	☒	On	☐	1459.53	2000.0	☐	28.00	27.93	++++
	Parallel	☒	On	☐	6.35	1000.0	☐	40.00	38.78	+++
	Serial	☒	On	☐	486.71		☐	1400.0	1398.1	++++
Heart	Parallel	☒	On	☐	0.02	800.0	☐	45.00	30.72	
	Serial	☒	On	☐	0.01		☐	2000.0	1234.4	
SC	Maximum Dose	☒	On	☐	0.01			4000.0	2549.6	
Oropharynx	Serial	☒	On	☐	0.01		☐	2500.0	2231.8	
3cm	Serial	☒	On	☐	71.22		☐	4050.0	4042.4	+++
	Quadratic Overdose	☒	On	☐	0.01	4700.0	☐	100.0	76.3	
skin	Maximum Dose	☒	On	☐	665.58			5400.0	5418.5	++++
	Maximum Dose	☒	On	☐	9.19			4700.0	4694.9	+++
	Conformality	☒	On	☐	0.01		☐	0.70	0.27	




剂量数据及分布






IMRT乳腺癌术后调强计划设计总结




未使用Auto Flash



使用2CM Auto Flash



VMAT段弧设计





VMAT参数配置

Sequencing Parameters: VMAT

☒ Segment Shape Optimization

☒ Pilot Beamlets

Max Number of Arcs:

Max. # of Control Points Per Arc:

Target Dose Rate (MU/min):

Min. Segment Width (cm):

Fluence Smoothing:

☐ Constant Dose Rate

OK Cancel



VMAT函数配置

IMRT Constraints										
↑ ↓ Pareto Constrained IMRT Parameters Constraints Sensitivities										
Structure	Cost Function	Enabled	Status	Manual	Weight	Reference Dose (cGy)	Multicriterial	Isoconstraint	Isoeffect	Relative Impact
PTV50	Quadratic Overdose	☒	On	☐	0.01	5150.0		100.0	11.1	
	Underdose DVH	☒	Infeasible	☐	100.00	5000.0		95.00	97.80	++++
	Target EUD	☒	On	☐	1.00			5000.0	4806.4	
	Maximum Dose	☒	On	☐	1170.69			5350.0	5355.9	++++
	Target Penalty	☒	On	☐	1.00			5100.0	4969.6	
lung	Maximum Dose	☒	On	☐	0.01			5000.0	5000.4	
	Parallel	☒	On	☐	678.33	2000.0	☐	28.50	28.57	++++
	Parallel	☒	On	☐	0.04	500.0	☐	55.00	48.50	+
	Serial	☒	On	☐	288.60		☐	1400.0	1422.5	++++
Heart	Parallel	☒	On	☐	0.01	800.0	☐	45.00	23.74	
	Serial	☒	On	☐	0.06		☐	800.0	699.2	+
SC	Maximum Dose	☒	On	☐	0.01			4200.0	3310.9	
Oropharynx	Parallel	☒	On	☐	0.02	3500.0	☐	30.00	28.36	+
skin	Maximum Dose	☒	On	☐	111.90			5400.0	5380.4	++++
	Maximum Dose	☒	On	☐	8.00			4600.0	4605.8	+++
	Conformality	☒	On	☐	0.01		☐	0.70	0.21	
<click to add a new structure>										



VMAT剂量数据及分布





IMRT-VMAT计划跳数分析

Prescription

Prescription

Segments

Add Rx

Delete Rx

	Rx ID	Rx Site	Prescribe To			Rx Dose (cGy)	Number of Fractions	Fractional Dose (cGy)				
 Physician's Intent	A		Plan Isocenter	X	7.88	Y	-115.60	Z	35.51	5000.0	25	200.0

Actual Dose = 4657.6 cGy

Rescale

5000.0 cGy

to...

Weight beams by: ☒ Dose ☐ MU

Equal Weights

Beam	Description	Field ID	%	Lock	MU / Fx
1	v280			☐	362.55
2	v110			☐	306.84

Total MU / Fx

669.39

Rescale

5000.0 cGy

to...

Weight beams by: ☒ Dose ☐ MU

Equal Weights

Beam	Description	Field ID	%	Lock	MU / Fx
1	A310			☐	119.11
2	A330			☐	93.80
3	A290			☐	156.68
4	A130			☐	104.00
5	A150			☐	90.77
6	A110			☐	126.92

Total MU / Fx

691.28

THANK YOU!



15541260977



7470298@qq.com



7470298